

SI-TRACEABLE SYSTEM DEVELOPMENT (WP-2220)

LUNAR IRRADIANCE MEASUREMENTS WITH A PRECISION FILTER RADIOMETER

Natalia Kouremeti, Julian Gröbner , Stelios Kazadzis

Physikalisch-Meteorologisches Observatorium Davos, World Radiation Center, PMOD/WRC

Saulius Nevas, Peter Schneider, Kerstin Schwind

Physikalisch-Technische Bundesanstalt (PTB), Braunschweig and Berlin, Germany






Lunar Precision Filter Radiometers

- **Filter radiometer with 4 channels in a grid**

Lunar version: 675 nm, 412 nm, 500 nm, 862 nm (FWHM: ~ 5 nm)

Optimized for Irradiance Measurements

- **Temperature stabilized photodiodes**
- **Reference Plane: the precision aperture**
- **FOV : 1.2° plateau , 0.7° slope angle, homogeneity in plateau > 99.5%**
- **22-bit data acquisition system (SACRAM) specifically designed for the PFR.**




Lunar-PFR PFR-L-002 has been characterized at PTB, Braunschweig, Germany and provides irradiance measurements with an uncertainty of 0.3% at 412 nm, 500 nm, 675 nm and 862 nm with the framework of 19ENV04 MAPP.

Tunable Lasers In Photometry (TULIP) setup


- ps-OPO system
- Homogenized beam
- Reference detector: 3-element trap detector and equipped with a calibrated aperture, giving an uncertainty better than 0.1 %
- Wavelength scale: LSA
- Fully automated system

Characterization Measurements

- Spectral responsivity (s)
- Reference plane
- PFR Gain



PFR-L Characterization



Spectral responsivity uncertainty < 0.3%

TULIP - 2021

λ (nm)	s (mV/(W/m ²))	U_s (%,k=2)
861.75	12.96	0.26
501.39	9.78	0.25
411.95	10.88	0.27
675.39	6.80	0.18

- : u_{trap}
- : u_{stab}
- : u_{dist}
- : $u_{\text{PFR,dark}}$
- : u_{wl}
- : u_{aperture}
- : u_{homog}
- : u_{current}
- - - : u_s
- : s

Gain uncertainty 0.3%


Gain	TULIP 2021 $U=0.3\%$
Laboratory: 0	1.0
1	934.6
2	4451.4
Lunar: 3	25164.0

PFR-L Characterisation : ATLAS (Pulsed) vs TULIP

Gain uncertainty reduced by a factor of 3

	TULIP 2021	ATLAS 2017	ATLAS 2019
Gain	U=0.3%	U=2.5%	U=1.5%
Laboratory			
: 0	1.0	0.0	0.0
1	934.6	-1.8	-0.2
2	4451.4	-1.9	-0.2
Lunar: 3	25164.0	-1.8	0.1

Spectral Responsivity: the spectral characteristics of **The interference filters** highly depend on the measuring geometry.




Relative responsivity of 412 nm channel retrieved from measurements at TULIP and ATLAS setups.

highest discrepancies

Spectral responsivity uncertainty reduced from 1.5% to 0.3%

TULIP - 2021			Differences to TULIP ATLAS-2020	
λ (nm)	s (W/m ²)	U (%,k=2)	$\delta\lambda$ (nm)	δs (%)
861.75	12.96	0.26	0.2	-1.4
501.39	9.78	0.25	1.1	0.8
411.95	10.88	0.27	0.8	5.1
675.39	6.80	0.18	0.1	-2.5

PFR-L Characterisation - Comparison of Calibration Methods



Irradiance Standard vs Monochromatic Irradiance

An irradiance calibration was performed at PTB after the TULIP calibration using 200 W lamps. The 2 calibration methods gave equivalent results, well within their uncertainties.


channel	862 nm	500 nm	412 nm	675 nm
1000 mm	-0.30%	-0.40%	-0.90%	0.30%
1500 mm	-0.10%	0.20%	0.30%	0.20%

PFR-L Characterisation - Stability 2015-2021

Direct Irradiance Calibration Setup



- Reference irradiance source (1000W FEL-type lamp) calibrated at PTB.
- Motorized XYZ linear translators
- Motorized Rotation stages for azimuth and zenith angles.





Comparison of lamp calibrations performed since 2015 to TULIP calibration.

The gray lines: uncertainty of the lamp calibration.


Top-Of-Atmosphere Lunar Irradiance

$$\ln \left(\frac{I_{pfr}(\lambda)}{I_{RIMO}(\lambda)} \right) + t_{ray}m + t_{O_3}m_{O_3} = -mt_{aod}$$

Successful Langley Example:




Top-Of-Atmosphere Lunar Irradiance




	Difference LunarPFR-RIMO TOA Lunar Irradiance (%)				Combined calibration and regression uncertainty (% , k=2)				Lunar Phase
	862 nm	500 nm	412 nm	675 nm	862 nm	500 nm	412 nm	675 nm	
Individual Langley retrievals	7.47	9.47	7.80	10.27	0.33	0.35	0.50	0.29	-47
	7.09	9.16	7.69	10.03	0.33	0.35	0.51	0.28	-33
	6.46	9.79	7.92	9.91	0.33	0.33	0.37	0.27	-19
	6.64	9.97	7.79	10.25	0.33	0.32	0.35	0.27	-6
	6.16	9.22	8.00	10.07	0.35	0.34	0.37	0.29	58
Mean TOA difference (%)	6.76	9.52	7.84	10.10	0.40	0.40	0.47	0.36	
standard deviation (%)	0.52	0.36	0.12	0.15					
Combined expanded uncertainty of TOA Lunar irradiance of PFR (k=2, %)	1.60	1.41	1.27	1.25					

Top-Of-Atmosphere Lunar Irradiance

Mean TOA difference (%)	6.76	9.52	7.84	10.10
Difference WEHRLI -TSIS-1 (%)	1.52	-2.17	-2.24	-0.24



SI-Traceable AOD retrieval for Sun-PFR




Achieved:

- ✓ **Calibrations with $U < 0.5\%$**
- ✓ **ToA consistent with calibration $< 1\%$**

⇒ $\Delta_{AOD} \sim 0.005 \pm 0.005$



Comparison Campaign against the WMO AOD reference – FRC-V(Davos,2021)



- Lunar PFR characterized and can provide lunar irradiance with an expanded relative uncertainty $U < 0.5\%$
- The SI-traceable AOD retrievals for the characterized Sun-PFR and TSIS-1, QASUME-FTS TOA solar spectra provides AOD equivalent to the standard Langley calibration.
- The lunar irradiance phase variation from ROLO/RIMO seems to be predicted well with an uncertainty of less 1% ($k=2$) (within the lunar $\pm 50^\circ$ phase).
- It is essential to use lunar reflectivity to use state-of-the-art solar spectra (TSIS-1)


Next steps ...

- Analyze Davos data
- Increase our reference data set.
- Development of a second Lunar-PFR to expand the information to 450
- Organize field campaigns for solar and lunar measurements



Top-Of-Atmosphere Lunar Irradiance

- The lunar irradiance seems to be underestimated by 10%
 - ROLO - old relative responsivity of PFR (Tom Stone)
 - RIMO - resolution 1 nm, generic relative responsivity
- Differences on the convolved solar spectrum due to responsivity: ~10%



SI-traceable system development (WP-2220) Lunar Irradiance measurements with a PFR

- Spectral responsivity calibration of Solar/Lunar Precision Filter Radiometer and Precision Solar Spectroradiometer traceable to the SI.
- Measurements of solar & Lunar spectral irradiances with PFR and spectral solar irradiance with PSR at PMOD/WRC and retrieval of spectral AOD.
- Field campaign at pristine measurement site for validation of solar & Lunar spectral irradiance measurements from Solar/Lunar PFR. Location will be selected during the first half of the project.




Outputs:

- D.2.2.2-1 Report on the laboratory calibration of PFR and PSR.
- D.2.2.2-2 Dataset of solar & lunar spectral irradiance and AOD from Precision filter radiometer and PSR during the phase 2 period of the project.
- D.2.2.2-3 Dataset of spectral solar/lunar irradiance from field campaign with solar/lunar Precision Filter Radiometer.
- D.2.2.2-4 Report on the field campaign for the validation of solar & lunar spectral irradiance measurements with the Solar/Lunar PFR.


WP1 Task 1.3: Field of view properties

A1.3.3: FOV measurements using the MOON

Lunar – PFR-L-109



ch1	: Cntr=0.04 °, Plateau=1.2 °, U=1.41% (>0.99 of max) FWHM=2.45 [-1.23,1.22]
ch2	: Cntr=-0.05 °, Plateau=0.9 °, U=0.60% (>0.99 of max) FWHM=2.41 [-1.25,1.17]
ch3	: Cntr=-0.26 °, Plateau=0.6 °, U=0.31% (>0.99 of max) FWHM=2.40 [-1.25,1.15]
ch4	: Cntr=-0.06 °, Plateau=1.3 °, U=0.99% (>0.99 of max) FWHM=2.42 [-1.24,1.18]
I0 used	



Precision Filter Radiometers **description**

- **Filter radiometer with 4 channels in a grid**
Interference filters :

Sun version: 368 nm, 412 nm , 500 nm, 862 nm

Lunar version: 675 nm, 412 nm , 500 nm, 862 nm

FWHM: ~5 nm

Optimized for Irradiance Measurements

- **Temperature stabilized photodiodes**
- **Reference Plane: the precision aperture**
- **FOV : 1.2° plateau , 0.7° slope angle,**
homogeneity in plateau > 99.5%
- The PFR signal (V) is provided by a 22-bit data acquisition system (SACRAM) specifically designed for the PFR.
- SACRAM Linearity checked against a reference source calibrated at Metas

